

JUN 22 2009

VERIFICATION OF CALIBRATION REPORT

*Of DataMaster cdm Breath test Instrument
State of Alaska
Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program*

DataMaster cdm S/N: 130126 ✓		
Supervisor/Operator Performing the Verification Procedure:		
Name: Lt. Eric Saffell	ID#: 2853	Date: 6/17/09 A
Agency: Anchorage Airport Police		Phone #: 266-2579
Instrument Location: Anchorage Airport Police Department		
B	Alco S/N: X124656 ✓	Target Value: .079 ✓
	High Pressure: 950 psi	
	Alco Test Value Average: .078 ✓ 1 st Alco	.078 ✓ 2 nd Alco
Signature: <u>LT Eric Saffell</u> ✓		
(OVER)		
8MB 6/25/09		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of June 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: 130126 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126

JUNE 17, 2009 ✓

OPERATOR'S NAME:

SAFFELL/ERIC

OPERATOR'S NUMBER: 2853

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: ANCH

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124656 ✓

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.78 in ✓
ALCO TARGET .078 21:35
BLANK TEST .000 21:36
INTERNAL STANDARD VERIFIED 21:36
ALCO TU 29.78 in .078 21:36
BLANK TEST .000 21:37
SUBJECT SAMPLE .000 21:38
BLANK TEST .000 21:39
ALCO TU 29.78 in .078 21:39
BLANK TEST .000 21:40

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130126 ✓

JUNE 17, 2009 ✓

TIME 21:31 ✓

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.78 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

BT6 9/95